


MAINTENANCE OF CONTRA-ANGLES



Refer to the product's instructions for use for the full cleaning and sterilization maintenance procedures. This medical device is supplied in "non-sterile" condition. Please ensure that this procedure is followed before the device is used.

For implantology and surgery handpieces, please refer to the dedicated instructions for use.

GENERAL CAUTION

⊘ Do not immerse the instrument in a NaCl bath or in an ultrasonic cleaner.


1 PREPARATION



A. REMOVING THE BUR

- Remove the rotary bur from the chuck mechanism.

2 PRELIMINARY CLEANING



A. RINSING THE DEVICE

- Clean the exterior and the interior of the device under tap water.



B. REMOVING DIRT/DEPOSITS

- Dry the nozzles then the rest of the instrument with a sterile non-woven compress (low linting textiles).

3 DISINFECTION



A. CLEANING AND DISINFECTION

- Dip the device in a bath containing a cleaning product/disinfectant.
- Brush the device with a smooth, flexible brush. DO NOT USE a wire brush.
- Rinse the device twice under running tap water.



B. OPTIONAL: AUTOMATIC DISINFECTION

- Carry out automatic washing or cleaning/disinfection.**
- After cleaning/disinfection, check that there are no detergent residues and, if required, rinse again.
- After cleaning/disinfection, check that the device is dry. If required, manually dry the device (see 4).

***For more information, please refer to the manufacturer instructions for use*

MAINTENANCE


OF CONTRA-ANGLES

4 MANUAL CLEANING AND DRYING



C. INTERNAL AND EXTERNAL CLEANING AND DRYING WITH SPRAYNET

- Shake the **Spraynet** before use.
- Select the appropriate **Spraynet** nozzle ①*
- Keep the tip of the instrument pointing down and protect it with a sterile non-woven cloth.
- Activate the **Spraynet**.



- Dry the exterior of the contra-angle with a cloth soaked in **Spraynet**.

⚠ Always dry the device.


5 LUBRICATION



A. LUBRICATION WITH LUBRIFLUID

- Shake the **Lubrifluid** before use.
- Select the **Lubrifluid** nozzle ①*
- Keep the tip of the instrument pointing down and protect it with an absorbent cloth.
- Activate the **Lubrifluid**.

6 STERILIZATION



B. PLACE THE CONTRA-ANGLE IN A STERILIZATION BAG

C. STERILIZE IN ACCORDANCE WITH THE MANUFACTURER'S INSTRUCTIONS

- Only instruments bearing the  symbol can be sterilized in an autoclave.

⚠ Only sterilize instruments which have been properly cleaned, dried and lubricated.

*All the attachments are shown on the last page of this document.

You can download the full instructions from: <https://dental.bienair.com/ifu>

MAINTENANCE

ACCESSORIES



SPRAYNET
500 ml cleaning spray
Pack of 6
REF. 1600036-006



LUBRIFLUID
500 ml lubricant
Pack of 6
REF. 1600064-006



LUBRICARE 2
Maintenance system
REF. 1601156-001